

KENYA DAIRY BOARD

FORM OF ADENDA

TO ALL TENDERERS ON:

TENDER DOCUMENT FOR SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF LABORATORY EQUIPMENTS AT KENYA DAIRY BOARD NATIONAL LABORATORY AT UPPER KABETE.

REF: TENDER NO. KDB/G/109/15602

ADDENDUM NO. 1

DATE: 9th December, 2020

The attention of all Tenderers is called to the following Addendum. This Addendum is hereby included in and made a part of the tender Documents, whether or not attached thereto.

All requirements of the Original Specifications shall remain in force except as amended by this Addendum.

CLARIFICATIONS HAVE BEEN SOUGHT AS FOLLOWS:

- ITEM 1: VACUUM OVEN

Query 1: Kindly clarify whether you intend to use the vacuum oven for flammable solvents

Answer: The vacuum oven will not be used for flammable Solvents.

Query 2: Kindly clarify if you require the vacuum oven to be supplied with the pump.

Answer: Yes, the vacuum oven should be 65 liters capacity and should be supplied complete with a vacuum pump.

- ITEM 7. HIGH TEMPERATURE DATA LOGGER.

Query 3: You have given 3 different probe dimensions, kindly specify your required probe dimensions in terms of inches.

Answer: The probe dimensions should be 5.25 in X 0.188 in Diameter (133mm x 4.8mm dia)

- ITEM 12. FLUID VISCOMETER

Query 3: Kindly specify the speed range that you require.

Answer: The speed should be 200 r.p.m \pm 1r.p.m

Query 4: Kindly specify the viscosity range that you need.

Answer: Viscosity Range in Absolute Viscosity (27-5274 Cp)

- ITEM TOP PAN BALANCE 100MG-220MG

Query 5: Does this description of Top Pan Balance 100mg-220mg on page 32 form part of the specification.

Answer: Use Technical Specifications as specified on page 40

END OF ADDENDUM

I confirm I have received ADDENDUM NO. 1- KDB/G/109/15602

NAME _____

ORGANISATION _____

SIGN _____

DATE _____

Form Ref: F/4/34

Issue:2

Revision:0